



Evaluation of CDCOM version 1.6





Problem description

- CDCOM version 1.5.2:
As a first approximation of the position of the gold discs the gridlines of the phantom are used
- Variations in gridline position led to an inaccurate estimation.
- Combined with the size of the search area this led to incorrect scores for some specific phantom images.

Problem description

Example of a phantom image with incorrect score in CDCOM debug mode:






CDCOM version 1.6

- CDCOM version 1.6.0:

A template of the phantom with all disc positions is translated, rotated and skewed for an optimal fit using the clearly visible discs on the images.

- For this estimate of the position of the discs a relatively large search area is used
 - For the actual detection a similar search area compared to version 1.5.2. is used to get similar detection results
- 



Evaluation of CDCOM 1.6

- Internally two new CDCOM versions have been made with different methods to estimate the position of the gold discs
- This data was compared to CDCOM version 1.5.2
- Debug CDCOM versions have been made in which each step in the programme could be checked







Evaluation of CDCOM 1.6

- A CDMAM phantom was made with only thick discs
- Using subsets of the visible discs the methods to determine the position of the search area could be evaluated
- The three CDCOM versions (1.5.2 and two new versions) were compared to each other using numerous CDMAM series
- Partly in debug mode







Evaluation of CDCOM 1.6

- Series of images from **each** manufacturer have been used to evaluate CDCOM 1.6 (between 8 to 164 images per series)
 - Data from 13 phantom comparisons have been used (each of these comparisons incorporated 2 to 26 phantoms)
 - In total images from approximately 90 CDMAM phantoms have been obtained and used in the evaluation
- 
- 


Evaluation of CDCOM 1.6




Evaluation of CDCOM 1.6




Evaluation of CDCOM 1.6




Evaluation of CDCOM 1.6




Evaluation of CDCOM 1.6



Evaluation of CDCOM 1.6



Evaluation of CDCOM 1.6






Evaluation of CDCOM version 1.6

- CDCOM has been evaluated extensively using data from all available systems and phantom comparisons from EUREF, Artinis and several third parties
- The search area is in the correct position in CDCOM version 1.6



- 
- CDCOM version 1.6 was released as beta-version in September 2011
 - Potential feedback would be used to improve the programme
 - No feedback with incorrect scores or other bugs have been reported
 - Since January 1st 2012: CDCOM version 1.6-beta has been released as 'official' version (version 1.6)
- 